

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Fluconazole FLU 25 µg	092421140	2024.09.23

Ref.

9166 – 9166/1 – 9166/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	28-39 28
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	22-33 23
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	6
<i>Candida tropicalis</i> ATCC® 750	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	26-37 28

Batch Release

Approved

Date

30.09.2021

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.