

# Liofilchem®

## Certificate of Analysis

Product	Batch	Expiration date
Voriconazole VO 1 µg	071421052	2024.07.13

**Ref.**

9168 – 9168/1 – 9168/2

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)	
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	31-42	31
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	28-37	32
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	16-25	23

**Batch Release**

Approved

**Date**

23.07.2021

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.