

Quality control certificate

GC Agar Base

N° Cat: 1106

N° Lote: 103032

Re-test Date: 3/2024



Formula in g/l

Bacteriological agar	10	Dipotassium phosphate	4
Maize starch	1	Monopotassium phosphate	1
Peptone mixture	15	Sodium chloride	5

Quality control

Solubility	Appearance	Color of the dehydrated medium	Color of the prepared medium	Final pH (25°)	Lote pH
Opalescent	Fine powder	Beige	Amber, slightly opalescent	7,2±0,2	7,1

Microbiological test

Incubation conditions: (35±2 °C / 5-10 % CO2 atmosphere / 40-48 h).

Microorganisms	Specification
Neisseria meningitidis ATCC 13090	Good growth
Haemophilus influenzae ATCC 19418	Good growth
Neisseria gonorrhoeae ATCC 19424	Good growth
Streptococcus pyogenes ATCC 19615	Good growth
Streptococcus pneumoniae ATCC 6305	Good growth

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.