

Quality control certificate



AST Medium

N° Cat: 1481

N° Lote: 909173

Re-test Date: 9/2023

Formula in g/l

Dextrose	10	Bacteriological agar	20
Disodium phosphate	1	Sodium citrate	10
Tryptone	9	Yeast extract	5
Sodium phosphate	1		

Quality control

Solubility	Appearance	Color of the dehydrated medium	Color of the prepared medium	Final pH (25°)	Lote pH
w/o rest	Fine powder	Beige	Amber	6,6±0,2	6,6

Microbiological test

Incubation conditions: (32 °C / 24-48 h).

Microrganisms	Specification
Candida albicans ATCC 10231	Good growth

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.