

Quality control certificate



Blood Agar Base N° 3

N° Cat: 1494

N° Batch: 202113

Re-test Date: 2/2025

Formula in g/l

Bacteriological agar	10	Dipotassium phosphate	1
Gelatin	4	Monopotassium phosphate	4
Peptone	10,5	Starch	3
Heart infusion	10,4	Aqua calcis	0,08

Quality control

Solubility	Appearance	Color of the dehydrated medium	Color of the prepared medium	Final pH (25°)	Batch pH
May present a slight precipitate	Fine powder	Beige	Amber w/o blood. Cherry red with blood	7,3±0,2	7,2

Microbiological test

Incubation conditions: (35±2 °C / 24-48 h).

Microorganisms	Specification
Neisseria meningitidis ATCC 13090	Good growth
Campylobacter jejuni ATCC 29428	Good growth

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.