


### CERTIFICATE OF ANALYSIS

<b>Product Name: Prolex™ Streptococcal Grouping Latex Kit</b>		<b>Lot Number: C16095</b>
<b>Product Code: PL.030</b>		<b>Expiration Date: 2019-08-31</b>
<b>Product Code</b>	<b>Quantity</b>	<b>Components</b>
PL.031	1 x 3.0 ml	Streptococcal Group A Latex Reagent
PL.032	1 x 3.0 ml	Streptococcal Group B Latex Reagent
PL.033	1 x 3.0 ml	Streptococcal Group C Latex Reagent
PL.034	1 x 3.0 ml	Streptococcal Group D Latex Reagent
PL.035	1 x 3.0 ml	Streptococcal Group F Latex Reagent
PL.036	1 x 3.0 ml	Streptococcal Group G Latex Reagent
PL.037	1 x 3.2 ml	Streptococcal Extraction Reagent 1
PL.038	1 x 3.2 ml	Streptococcal Extraction Reagent 2
PL.039	2 x 8.0 ml	Streptococcal Extraction Reagent 3
PL.040	1 x 2.0 ml	Polyvalent Positive Control (A,B,C,D,F,G)
	300	Mixing Sticks
PL.092-48	48 cards	Test Cards
<b>Additional Information</b>		
Storage Conditions:	2-8°C	
Precautions:	Refer to Instructions for Use	
<b>Quality Control Testing</b>		
	<b>Criteria</b>	<b>Result</b>
	Chemical extraction of Strep group specific antigens.	PASS
	Strong agglutination of the test isolate when tested against the corresponding latex reagent.	PASS
	No agglutination reaction of the test isolate when tested against heterologous latex reagents.	PASS
	Strong agglutination of the Positive Control when tested against the latex reagents.	PASS
<b>Release</b>		
2018 05 25		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206  
20 Mural Street, Unit #4, Richmond Hill, Ontario,  
L4B 1K3  
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
21 Cypress Blvd., Suite 1070, Round Rock,  
Texas 78665-1034  
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
uksupport@pro-lab.com