

CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping Reagent Latex Suspension Group A	Lot Number: C16301D
Product Code: PL.031	Expiration Date: 2020-01-31


Additional Information

Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

Appearance	Reaction	Result
Homogenous suspension with no agglutination	N/A	Satisfactory
Organisms		
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Positive	PASS
<i>Streptococcus</i> Group A - Clinical Isolate	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Negative	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Negative	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2)	Negative	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Negative	PASS
<i>Listeria monocytogenes</i> NCTC 11994	Negative	PASS
Control		
Polyvalent Positive Control (PL.040)	Positive	PASS

Release

2018 08 28	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com