

### CERTIFICATE OF ANALYSIS

|  |                                    |  |
|--|------------------------------------|--|
| <b>Product Name:</b> Streptococcal Grouping Reagent Latex Suspension Group A | <b>Lot Number:</b> C19149C         |  |
| <b>Product Code:</b> PL.031  | <b>Expiration Date:</b> 2022-09-30 |  |


#### Additional Information

|                     |                               |
|---------------------|-------------------------------|
| Storage Conditions: | 2-8°C                         |
| Precautions:        | Refer to Instructions for Use |

#### Quality Control Testing

| Appearance  | Reaction | Result       |
|---|----------|--------------|
| Homogenous suspension with no agglutination           | N/A      | Satisfactory |
| <b>Organisms</b>                                      |          |              |
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A)    | Positive | PASS         |
| <i>Streptococcus</i> Group A - Clinical Isolate       | Positive | PASS         |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B)  | Negative | PASS         |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Negative | PASS         |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D)     | Negative | PASS         |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2) | Negative | PASS         |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G)         | Negative | PASS         |
| <i>Listeria monocytogenes</i> NCTC 11994              | Negative | PASS         |
| <b>Control</b>  |          |              |
| Polyvalent Positive Control (PL.040)                  | Positive | PASS         |

#### Release

|            |  |
|------------|--|
| 2021 05 25 |  |
| Date       | Quality Control  |

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424  
 21 Cypress Blvd., Suite 1155, Round Rock,  
 Texas 78665-1046  
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.com