

CERTIFICATE OF ANALYSIS

| | | |
|--|------------------------------------|--|
| Product Name: Streptococcal Grouping Reagent Latex Suspension Group D | Lot Number: C17155A | |
| Product Code: PL.034 | Expiration Date: 2020-10-31 | |


Additional Information

| | |
|---------------------|-------------------------------|
| Storage Conditions: | 2-8°C |
| Precautions: | Refer to Instructions for Use |

Quality Control Testing

| Appearance | Reaction | Result |
|---|----------|--------------|
| Homogenous suspension with no agglutination | N/A | Satisfactory |
| Organisms | | |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D) | Positive | PASS |
| <i>Enterococcus faecalis</i> ATCC 49452 (Group D) | Positive | PASS |
| <i>Enterococcus faecalis</i> ATCC 29212 (Group D) | Positive | PASS |
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A) | Negative | PASS |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B) | Negative | PASS |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Negative | PASS |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2) | Negative | PASS |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G) | Negative | PASS |
| <i>Listeria monocytogenes</i> NCTC 11994 | Negative | PASS |
| Control | | |
| Polyvalent Positive Control (PL.040) | Positive | PASS |

Release

| | |
|------------|--|
| 2019 06 06 |  |
| Date | Quality Control |

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com