

CERTIFICATE OF ANALYSIS

| | | |
|--------------------------------------------------------------------|------------------------------------|--|
| Product Name: Streptococcal Grouping - Extraction Reagent 2 | Lot Number: C16215G | |
| Product Code: PL.038 | Expiration Date: 2020-06-30 | |

Additional Information


| | |
|---------------------|-------------------------------|
| Storage Conditions: | 2-8°C |
| Precautions: | Refer to Instructions for Use |

Quality Control Testing

| Organism | Reaction* | Result |
|-------------------------------------------------------|-----------|--------|
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A) | Positive | PASS |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B) | Positive | PASS |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Positive | PASS |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D) | Positive | PASS |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F) | Positive | PASS |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G) | Positive | PASS |

*When used in conjunction with Extraction Reagents 1 and 3 to extract group specific carbohydrate antigens and tested against homologous latex reagents.

Release

| | |
|------------|---------------------------------------------------------------------------------------|
| 2018 08 24 |  |
| Date | Quality Control |

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com