


### CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Streptococcal Grouping Reagent Polyvalent Positive Control	<b>Lot Number:</b> C11312B	
<b>Product Code:</b> PL.040	<b>Expiration Date:</b> 2015 07	

Additional Information	
Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing		
Grouping Reagent	Reaction	Result
Streptococcal Grouping Reagent Latex Suspension Group A (PL.031)	Positive	PASS
Streptococcal Grouping Reagent Latex Suspension Group B (PL.032)	Positive	PASS
Streptococcal Grouping Reagent Latex Suspension Group C (PL.033)	Positive	PASS
Streptococcal Grouping Reagent Latex Suspension Group D (PL.034)	Positive	PASS
Streptococcal Grouping Reagent Latex Suspension Group F (PL.035)	Positive	PASS
Streptococcal Grouping Reagent Latex Suspension Group G (PL.036)	Positive	PASS

Release	
2014 02 10	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
 21 Cypress Blvd., Suite 1070, Round Rock,  
 Texas 78665-1034  
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.com