


CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping - Prolex™ Extraction Reagent Set		Lot Number: C11404
Product Code: PL.046		Expiration Date: 2016 01
Product Code	Quantity	Components
PL.047	1 x 8.0 ml	Streptococcal Extraction Reagent 1
PL.048	1 x 8.0 ml	Streptococcal Extraction Reagent 2
PL.049	1 x 8.0 ml	Streptococcal Extraction Reagent 3
Additional Information		
Storage Conditions:		2-8°C
Precautions:		Refer to Instructions for Use
Quality Control Testing		
Organism	Reaction*	Result
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Positive	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Positive	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS
*When Extraction Reagents 1, 2 and 3 are used in conjunction to extract group specific carbohydrate antigens and tested against homologous latex reagents.		
Release		
2014 02 05		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
21 Cypress Blvd., Suite 1070, Round Rock,
Texas 78665-1034
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.com