


CERTIFICATE OF ANALYSIS

Product Name: Prolex™ <i>E. coli</i> O157 Latex Test Reagent Kit		Lot Number: C19782
Product Code: PL.070B		Expiration Date: 2023-03-31
Product Code	Quantity	Components
PL.072B	1 x 3.1 ml	<i>E. coli</i> O157 Latex Reagent
PL.074B	1 x 1.5 ml	<i>E. coli</i> O157 Positive Control
PL.077B	1 x 1.5 ml	<i>E. coli</i> O157 Latex Reagent Negative Control
	75	Mixing Sticks
PL.092-20	20 cards	Test Cards
Additional Information		
Storage Conditions:		2-8°C
Precautions:		Refer to Instructions for Use
Quality Control Testing		
Criteria		Result
Strong agglutination of Latex Reagent when tested against <i>E. coli</i> serotype O157 isolates.		PASS
No agglutination reaction of Latex Reagent when tested against non O157 serotypes <i>E. coli</i> , and <i>E. hermanii</i> isolates.		PASS
No agglutination reaction of Negative Control when tested against <i>E. coli</i> and <i>E. hermanii</i> isolates.		PASS
Strong agglutination of the <i>E. coli</i> O157 Positive Control when tested against Latex Reagent.		PASS
No agglutination of the <i>E. coli</i> O157 Positive Control when tested against Negative Control.		PASS
Release		
2021 09 10		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com