

CERTIFICATE OF ANALYSIS

Product Name:	Staph Latex Kit			Lot Number:	C12648	
Product Code:	PL.080B	Expiration Date:		2016 08		
Product Code	Quantity	Components				
PL.083B	2 x 2.5 ml		Latex Reagent			
PL.085B	1 x 2.5 ml	Negative C	Negative Control Latex Reagent			
	200	Mixing Sticks				
PL.092-20	20 cards	Test Cards	ls			
Additional Inform						
Storage Conditions: 2-8°C						
Precautions:	cautions: Refer to Instructions for Use					
Quality Control T	esting					
Criteria					Result	
Strong agglutination of Staph Latex Reagent when tested against Staphylococcus aureus.					PASS	
No agglutination of Negative Control when tested against test isolates.					PASS	
Release						
2015 05 07			Solly	Carg-		
Date				Quality Control		

Canada 🕾 800 268 2341 Fax 905 731 0206 20 Mural Street, Unit #4, Richmond Hill, Ontario, L4B 1K3 support@pro-lab.com U.S.A. 🕾 800 522 7740 Fax 800 332 0450 21 Cypress Blvd., Suite 1070, Round Rock, Texas 78665-1034 ussupport@pro-lab.com U.K. 🕾 0151 353 1613 Fax 0151 353 1614 3 Bassendale Road, Bromborough, Wirral, Merseyside CH62 3QL uksupport@pro-lab.com