

CERTIFICATE OF ANALYSIS

Product Name:				Lot Number:	C15609	
Product Code:	PL.080B	Expiration Date: 2019 03				
Product Code	Quantity	Compone	Components			
PL.083B	2 x 2.5 ml		Staph Latex Reagent			
PL.085B	1 x 2.5 ml	Negative C	egative Control Latex Reagent			
	200	Mixing Stic				
PL.092-20	20 cards	Test Cards	ds			
Additional Information						
Storage Conditions: 2-8°C						
Precautions:	Precautions: Refer to Instructions for Use					
Quality Control Testing						
Criteria					Result	
Strong agglutination of Staph Latex Reagent when tested against Staphylococcus aureus.					PASS	
No agglutination of Negative Control when tested against test isolates.					PASS	
Release						
2017 12 05			Soleka			
Date				Quality Control		