


CERTIFICATE OF ANALYSIS

Product Name: Staph Latex Kit		Lot Number: C18562
Product Code: PL.080B		Expiration Date: 2021-08-31
Product Code	Quantity	Components
PL.083B	2 x 2.5 ml	Staph Latex Reagent
PL.085B	1 x 2.5 ml	Negative Control Latex Reagent
	200	Mixing Sticks
PL.092-20	20 cards	Test Cards
Additional Information		
Storage Conditions:	2-8°C	
Precautions:	Refer to Instructions for Use	
Quality Control Testing		
	Criteria	Result
	Strong agglutination of Staph Latex Reagent when tested against <i>Staphylococcus aureus</i> .	PASS
	No agglutination of Negative Control when tested against test isolates.	PASS
Release		
2020 04 22		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com