


CERTIFICATE OF ANALYSIS

Product Name: Prolex™ <i>E.coli</i> Non-O157 Identification Kit		Lot Number: C16189
Product Code: PL.1070		Expiration Date: 2019 11
Product Code	Quantity	Components
PL.1071	1 x 2.5 ml	Prolex™ <i>E.coli</i> O26 Latex Reagent
PL.1072	1 x 2.5 ml	Prolex™ <i>E.coli</i> O45 Latex Reagent
PL.1073	1 x 2.5 ml	Prolex™ <i>E.coli</i> O103 Latex Reagent
PL.1074	1 x 2.5 ml	Prolex™ <i>E.coli</i> O111 Latex Reagent
PL.1075	1 x 2.5 ml	Prolex™ <i>E.coli</i> O121 Latex Reagent
PL.1076	1 x 2.5 ml	Prolex™ <i>E.coli</i> O145 Latex Reagent
PL.1077	1 x 2.5 ml	Prolex™ Negative Control Latex Reagent
	300	Mixing Sticks
	2 x 20 cards	Test Cards
Additional Information		
Storage Conditions:		2-8°C
Precautions:		Refer to Instructions for Use
Quality Control Testing		
	Criteria	Result
Strong agglutination of Latex Reagents when tested against homologous <i>E.coli</i> Non-O157 serotype isolates.		PASS
No agglutination reaction of Latex Reagents when tested against heterologous <i>E.coli</i> Non-O157 serotypes and <i>E.coli</i> O157 isolates.		PASS
No agglutination reaction of Negative Control when tested against <i>E.coli</i> Non-O157 serotype isolates.		PASS
Release		
2018 06 12		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com