

CERTIFICATE OF ANALYSIS

Product Name:	Prolex™ Staph	Xtra Latex Kit	Lot Number:	C12992
Product Code:	PL.1080		Expiration Date:	2017 03
Product Code	Quantity	Compone	nts	
PL.1083	2 x 2.5 ml	Staph Xtra	tra Latex Reagent	
PL.1085	1 x 5.0 ml	Staph Xtra Negative Control Latex Reagent		
	200	Mixing Sticks		
	20 cards	Test Cards		
Additional Inform	nation			
Storage Conditions: 2-8°C				
Precautions:	Refer to Instructions for Use			
Quality Control	Testing			
Criteria				Result
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.				
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds. PASS				
* MRSA - (Methic	illin-resistant Staph	ylococcus aureu	5)	
Release				
2015 08 18			2	12
Date			Quality Control	