

CERTIFICATE OF ANALYSIS

Product Name:	Prolex™ Staph Xtra Latex Kit Lo			C11711	
Product Code:	PL.1080		Expiration Date:	2015 12	
Product Code	Quantity	Compone	nts		
PL.1083	2 x 2.5 ml	Staph Xtra	a Latex Reagent		
PL.1085	1 x 5.0 ml	Staph Xtra	n Xtra Negative Control Latex Reagent		
	200	Mixing Stic	ks		
	20 cards	Test Cards			
Additional Inform	nation				
Storage Conditions: 2-8°C					
Precautions:	Refer to Instructions for Use				
Quality Control T	esting				
Criteria				Result	
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.					
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds. PASS					
* MRSA - (Methicillin-resistant <i>Staphylococcus aureus</i>)					
Release					
2014 05 28			Shilpa Tregos	l	
Date			Quality Control		