

## **CERTIFICATE OF ANALYSIS**

Product Name:	Prolex™ Staph Xtra Latex Kit Lot Numb			C14494	
Product Code:	PL.1080		Expiration Date:	2018 06	
Product Code	Quantity	Compone	nts		
PL.1083	2 x 2.5 ml	Staph Xtra	ra Latex Reagent		
PL.1085	1 x 5.0 ml	Staph Xtra Negative Control Latex Reagent			
	200	Mixing Stic	eks		
	20 cards	Test Cards	5		
Additional Inform					
Storage Conditions: 2-8°C					
Precautions:	Refer to Instructions for Use				
Quality Control Testing					
Criteria				Result	
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.				PASS	
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds.				PASS	
* MRSA - (Methicillin-resistant Staphylococcus aureus)					
Release					
2016 11 28			C. Le Coche		
Date			Quality Control		