



**CERTIFICATE OF ANALYSIS**

<b>Product Name:</b> Proflow H.pylori	<b>Lot Number:</b> P-273
<b>Product Code:</b> PL.3006	<b>Expiration Date:</b> 2022 07 31

**Additional Information**

Test conditions: As defined in the IFU.	Storage conditions: 2-30°C	The product is intended for in vitro use only.
-----------------------------------------	----------------------------	------------------------------------------------

**Quality Control Testing**

Functionality test results: visual interpretation at 10 minutes

1. Sensitivity test.  
Dilutions from standard: approved.

2. Specificity test.  
Materials prepared in product buffer: approved.

Quality Control result

Approved.

**Release**

2020/07/30	<i>mawr</i>
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
20 Mural Street, Unit #4, Richmond Hill, Ontario,  
L4B 1K3  
support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424  
21 Cypress Blvd., Suite 1155, Round Rock,  
Texas 78665-1046  
support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
uksupport@pro-lab.com