

CERTIFICATE OF ANALYSIS

Product Name: <i>E. coli</i> O157 Antiserum	Lot Number: C12699	
Product Code: PL.6250	Expiration Date: 2016 11	


Additional Information

Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

Organisms	Reaction	Result
<i>E. coli</i> O157:H7 ATCC 35150	Positive	PASS
<i>E. coli</i> O157:H7 ATCC 43895	Positive	PASS
<i>E. coli</i> O157:H- Clinical Isolates	Positive	PASS
<i>E. coli</i> non O157 serotype Clinical Isolates	Negative	PASS
<i>E. hermanii</i> Clinical Isolates	Negative	PASS

Release

2015 07 27	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
21 Cypress Blvd., Suite 1070, Round Rock,
Texas 78665-1034
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.com