


**CERTIFICATE OF ANALYSIS**

<b>Product Name:</b> <i>E. coli</i> O157 Antiserum	<b>Lot Number:</b> C18073A
<b>Product Code:</b> PL.6250	<b>Expiration Date:</b> 2021 05

<b>Additional Information</b>	
Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

<b>Quality Control Testing</b>		
<b>Organisms</b>	<b>Reaction</b>	<b>Result</b>
<i>E. coli</i> O157:H7 ATCC 35150	Positive	PASS
<i>E. coli</i> O157:H7 ATCC 43895	Positive	PASS
<i>E. coli</i> O157:H- Clinical Isolates	Positive	PASS
<i>E. coli</i> Non-O157 serotype Clinical Isolates	Negative	PASS
<i>E. hermanii</i> Clinical Isolates	Negative*	PASS
*Cross reactions may be present.		

<b>Release</b>	
2019 11 15	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
20 Mural Street, Unit #4, Richmond Hill, Ontario,  
L4B 1K3  
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
21 Cypress Blvd., Suite 1070, Round Rock,  
Texas 78665-1034  
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
3 Bassendale Road, Bromborough, Wirral,  
Merseyside CH62 3QL  
uksupport@pro-lab.com