



www.pro-lab.com

**CERTIFICATE OF ANALYSIS**

<b>Product Name:</b> CRYSTAL VIOLET	<b>Lot Number:</b> K29999
<b>Product Code:</b> PL.7000	<b>Expiration Date:</b> 2021 06 30

**Additional Information**

Description: Deep violet solution with slight odour.	Concentrate Lot number: K29976 Production date: 2020/06/23	Storage conditions: 15-25°C Precautions: refer to SDS
--	---	--

**Quality Control Testing**

Challenged with test organisms:  
*Escherichia coli* NCTC 10418  
 Oxford *Staphylococcus aureus* NCTC 6571  
 Haemolytic *Streptococcus* Group A NCTC 8198

Test results:  
 Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:  
 We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

**Release**

2020/06/23	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450  
 21 Cypress Blvd., Suite 1070, Round Rock,  
 Texas 78665-1034  
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.com