

CERTIFICATE OF ANALYSIS

Product Name: Mucolyse	Lot Number: C11127	
Product Code: PL.701	Expiration Date: 2015 10	

Additional Information


Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

	Criteria	Result
Appearance		
Pellet	Compact white pellet	PASS
Reconstituted Supplement	Clear colourless solution	PASS
Sterility Test		
	Conforms	PASS
Organism		
<i>Haemophilus influenzae</i> ATCC 10211	Growth at 24 hours* Good Growth	PASS
<i>Streptococcus pneumoniae</i> ATCC 6305	Good Growth	PASS

*Compared to control plates not containing supplement.

Release

2013 11 15	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
21 Cypress Blvd., Suite 1070, Round Rock,
Texas 78665-1034
ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.com