

CERTIFICATE OF ANALYSIS

Product Name: Mucolyse	Lot Number: C17083
Product Code: PL.701	Expiration Date: 2021-03-31

Additional Information

Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

	Criteria	Result
Appearance		
Pellet	Compact white pellet	PASS
Reconstituted Supplement	Clear colourless solution	PASS
Sterility Test		
	Conforms	PASS
Organism		
<i>Haemophilus influenzae</i> ATCC 10211	Growth at 24 hours* Good Growth	PASS
<i>Streptococcus pneumoniae</i> ATCC 6305	Good Growth	PASS

*Compared to control plates.

Release

2019 04 01	
Date	Quality Control