



www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name: SAFRANIN	Lot Number: K32101
Product Code: PL.7012	Expiration Date: 2022 07 31

Additional Information

Description: Deep red solution with slight odour.	Concentrate Lot number: K31989 Production date: 2021/08/03	Storage conditions: 15-25°C Precautions: refer to SDS
---	---	--

Quality Control Testing

Challenged with test organisms:
Escherichia coli NCTC 12241
 Oxford *Staphylococcus aureus* NCTC 12981

Test results:
 Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:
 We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

2021/08/09	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com