



www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name: LUGOLS IODINE	Lot Number: K18775
Product Code: PL.8010	Expiration Date: 2015 09

Additional Information

Description: Dark brown solution.	Concentrate Lot number: 6376 Production date: 2014/09/11	Storage conditions: 15-25°C Precautions: refer to MSDS
-----------------------------------	---	---

Quality Control Testing

Challenged with test organisms:
Escherichia coli NCTC 10418
 Oxford *Staphylococcus aureus* NCTC 6571
 Haemolytic *Streptococcus* Group A NCTC 8198

Test results:
 Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:
 We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

14/09/11	<i>maun .</i>
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com