

QUALITY CONTROL CERTIFICATE

AmpC Confirm ID Pack, Neo-Sensitabs

REF: 98007

LOT: 1206-1

Date of analysis: 11-2011, 01-2012, 03-2011, 11-2011

Date of expiry: 03-2014

Appearance: Light yellow, plane tablets, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

Staphylococcus aureus ATCC 25923

Pseudomonas aeruginosa ATCC 27853

E. coli FN9414

Ent. Cloaceae FN9475

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



Ole Skyggebjerg
Quality Assurance Manager