

QUALITY CONTROL CERTIFICATE

KPC, MBL and OXA-48 Kit, Neo-Sensitabs

REF: 98015

LOT: 1303-1

Date of analysis: 12-2012, 05-2012, 02-2012, 11-2012, 02-2013

Date of expiry: 02-2015

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> NCTC 13438	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 1705	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 2146	MBL Positive
<i>K. pneumoniae</i> NCTC 13439	MBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals
Quality Assurance Manager