

QUALITY CONTROL CERTIFICATE

Detection of hVISA, GISA, VRE, Daptomycin resistance

REF: 98017

LOT: 1301-1

Date of analysis: 11-2012, 09-2012, 05-2012

Date of expiry: 2015-05

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Test strains	Results
<i>Staph. aureus</i> ATCC 700698	hVISA and Daptomycin susceptibility
<i>Staph. aureus</i> AST 403	GISA and Daptomycin susceptibility
<i>Enterococcus faecalis</i> ATCC 51299	Van B positive
<i>Staph. aureus</i> AST 408	GISA and Daptomycin resistant

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals
Quality Assurance Manager