

QUALITY CONTROL CERTIFICATE

KPC/MBL in *P.aeruginosa*/Acinetobacter Confirm Kit, Neosensitabs

REF: 98020

LOT: 1401-1

Date of expiry: 2016-05

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains


<i>P.aeruginosa</i> ATTC 10145	MBL Positive
<i>P.aeruginosa</i> CCUG 59626	MBL Positive
<i>K. pneumoniae</i> CCUG 56233	KPC Positive
<i>K. pneumoniae</i> NCTC 13438	KPC Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals
Quality Assurance Manager