

## QUALITY CONTROL CERTIFICATE

### **KPC/MBL in *P.aeruginosa*/Acinetobacter Confirm Kit, Neosensitabs**

REF: 98025

LOT: 1701-1

Date of expiry: 2019-06

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains


<i>P.aeruginosa</i> ATTC 10145	MBL Positive
<i>P.aeruginosa</i> CCUG 59626	MBL Positive
<i>K. pneumoniae</i> CCUG 56233	KPC Positive
<i>K. pneumoniae</i> NCTC 13438	KPC Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Michela Venturi  
Manager Quality Assurance