

CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

Rev. 1.0 - 2015.07.01

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK16003	
MANUFACTURING DATE: 12/05/2016	EXPIRY DATE: 12/09/2017

◆ *Reconstituting and Diluting Solutions*

Eugonic broth

Batch N.:FLA16002

Expire date:
01/04/2018

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	OK	OK	34.5%	<5%

* Volume, transparency, colour

◆ *Lyophilized Microorganism*

Batch N: 15003

Expire date: 01/12/2017

	Growth inter batch (CV%≤35%)	Physical check**
Expected	Compliant	Compliant
Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2°C – 8 °C

Date of Approval:12/05/2016

Responsible for batch release


