

# CERTIFICATE OF ANALYSIS

## POSITIVE CONTROL KIT

Rev. 1.1 - 2016.06.22

<b>PRODUCT CODE: SI 190.911</b>	<b>NAME: POSITIVE CONTROL KIT</b>
<b>BATCH N.: PCK19002</b>	
<b>MANUFACTURING DATE: 19/02/2019</b>	<b>EXPIRY DATE: 19/04/2020</b>

◆ *Reconstituting and Diluting Solutions*

Eugonic broth

Batch N.: FLA19001

Expire date:  
07/02/2021

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
<b>Expected</b>	Compliant	No growth	CV%≤35%	CV%≤35%
<b>Found</b>	OK	OK	21,6%	<5%

\* Volume, transparency, colour

◆ *Lyophilized Microorganism*

Batch N: 18002

Expire date: 18/07/2020

	Growth inter batch (CV%≤35%)	Physical check**
<b>Expected</b>	Compliant	Compliant
<b>Found</b>	OK	OK

\*\* Aspect

**STORAGE CONDITIONS: 2°C – 8 °C**

**Date of Approval: 26/02/2019**

**Pividori Denis:**  
*Manufacturing Manager*


