



CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK20004	EXPIRY DATE: 14/07/2021
MANUFACTURING DATE: 20/05/2020	

◆ RECONSTITUTING AND DILUTING SOLUTIONS

Eugonic broth

Batch N.: FLA20003

Expire date: 28/04/2020

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	OK	OK	14,8%	<5%

* Volume, transparency, colour

◆ LYOPHILIZED MICROORGANISM

Batch N.: 20001

Expire date: 04/03/2022

	Growth inter batch (CV% ≤35%)	Physical check**
Expected	Compliant	Compliant
Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2 ÷ 8 °CDate of Approval:
22/05/2020Pividori Denis:
Manufacturing Manager

Date of Release:
22/05/2020

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released

Release authorized by Quality Assurance

ALIFAX S.r.l.

Registered and head office
via Petrarca, 2/1 - 35020 Polverara PD Italy
Tel. (+39) 049 0992000 Fax (+39) 049 5855434
info@alifax.com - info.alifax@pec.alifax.com

Entered in Padova companies register at the n. 04337640280. Company stock 10.000.000 entirely deposited.
Company with single partner and subjected to direction and coordination of Alifax Holding S.p.A.

R&D, Design and Manufacturing Dept
via Merano, 30,33045 Nimis UD Italy
Tel. (+39) 0432 197900 Fax (+39) 0432 547378
info.udine@alifax.com - info.udine@pec.alifax.com

www.alifax.com