

# **CERTIFICATE OF ANALYSIS**

## POSITIVE CONTROL KIT

PRODUCT CODE: SI 190.911 NAME: POSITIVE CONTROL KIT

BATCH N.: PCK21006 EXPIRY DATE: 16/09/2022

MANUFACTURING DATE: 16/07/2021

### **♦ RECONSTITUTING AND DILUTING SOLUTIONS**

Eugonic broth

Batch N.: FLA21002

Expire date: 08/06/2023

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	ОК	ОК	<b>2</b> 0,5%	17,7%

<sup>\*</sup> Volume, transparency, colour

#### **♦ LYOPHILIZED MICROORGANISM**

**Batch N.:** 21001

Expire date: 12/01/2023

	Growth inter batch (CV% ≤35%)	Physical check**
Expected	Compliant	Compliant
Found	ОК	ОК

<sup>\*\*</sup> Aspect

STORAGE CONDITIONS: 2 ÷ 8 °C

Date of Approval:

16/07/2021

Pividori Denis: Manufacturing Manager

Date of Release:

16/07/2021

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released

Release authorized by Quality Assurance

#### ALIFAX S.r.I.

Registered and head office via Petrarca, 2/1 - 35020 Polverara PD Italy Tel. (+39) 049 0992000 Fax (+39) 049 5855434 info@alifax.com - info.alifax@pec.alifax.com R&D, Design and Manufacturing Dept via Merano, 30,33045 Nimis UD Italy Tel. (+39) 0432 197900 Fax (+39) 0432 547378 info.udine@alifax.com - info.udine@pec.alifax.com

©Copyright 2020. This document is exclusive property of ALIFAX S.r.I, confidential and not freely disclosable. All rights reserved.