

## **CERTIFICATE OF ANALYSIS**

ESR\_COA\_SI305#2174A.pdf

Product Code	Name	Quantity manufactured	Type of Tubes
SI 305.100-A	LATEX CONTROLS	100	Greiner
SI 305.300-A	LATEX CONTROLS	158	Greiner

Lot Number	Manufacturing Date	Expiry Date	
2174A	11/03/2021	07/09/2021	

Standard Values ESR (mm/h)			F	ESR (mm/h) Measured Values					Internal	Volume			
	Transmittance	(min-Val-max)		(min-Val-max)		•	Instrument	Serial Number		OK	Aggregation	checked	
Level 2	1534	6	-	11	9	9	9	TEST 1 BCL		2173	$\square$	NO	YES
Level 3	980	15	-	22	19	20	20	TEST1 THL		T13290	$\square$	NO	YES
Level 4	637	56	-	74	66	66	68	ROLLER 20 PN		133	$\square$	NO	YES

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

Released By Quality ....... Falcone Francesco Date of Release 11/03/2021

This is a computer-generated document. No signature is required.

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



## **Manufacturing Site**

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com