

## **CERTIFICATE OF ANALYSIS**

## ESR\_COA\_SI305#2224A.pdf

Name	Quantity manufactured	Type of Tubes
LATEX CONTROLS	180	Greiner
LATEX CONTROLS	142	Greiner
	Name  LATEX CONTROLS  LATEX CONTROLS	LATEX CONTROLS manufactured 180

Lot Number	Manufacturing Date	Expiry Date	
2224A	01/09/2021	28/02/2022	

	Standard Values - ESR (mm/h)			FCD (mame/h)			Measured	Values			Internal	Volume	
	Transmittance		n-Val-n		ESR (mm/h) (min-Val-max)			Instrument		Serial Number	OK	Aggregation	checked
Level 2	1534	6	-	11	8	9	9	TEST 1 BCL		2173	X	NO	YES
Level 3	980	15	-	22	18	19	20	TEST1 THL		T13290	X	NO	YES
Level 4	637	56	-	<i>74</i>	63	66	68	ROLLER 20 PN		133	X	NO	YES

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

This is a computer-generated document. No signature is required.

**Registered Office** 

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



**Manufacturing Site** 

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com