



# CERTIFICATE OF ANALYSIS

ESR\_COA\_SI305#1840A.pdf

Product Code <b>SI 305.300-A</b>	Name <b>Latex Controls (30 Tests)</b>	Quantity manufactured <b>177</b>	Type of Tubes <b>Greiner</b>
-------------------------------------	--	-------------------------------------	---------------------------------

Lot Number <b>1840A</b>	Manufacturing Date <b>30/10/2017</b>	Expiry Date <b>28/04/2018</b>		
----------------------------	---	----------------------------------	--	--

Standard Values			Measured Values				Internal Aggregation	Volume checked		
Transmittance	ESR (mm/h) (min-Val-max)		ESR (mm/h) (min-Val-max)			Instrument	Serial Number	OK		
Level 2	1534	6 - 11	8	9	9	TEST 1 BCL	2173	<input checked="" type="checkbox"/>	NO	YES
Level 3	980	15 - 22	19	19	19	TEST1 THL	PRDX	<input checked="" type="checkbox"/>	NO	YES
Level 4	637	56 - 74	64	65	66	ROLLER 20 PN	133	<input checked="" type="checkbox"/>	NO	YES

Date of Control **30/10/2017**

Controlled By **Cirio Alberto**

Date of Approvation **30/10/2017**

Approved By **Pividori Denis**

**Reagents Production Manager**

## Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy  
Tel (+39) 049 0992000 - Fax (+39) 049 5855434  
Mail PEC: info.alifax@pec.alifax.com  
[info@alifax.com](mailto:info@alifax.com) - [www.alifax.com](http://www.alifax.com)



## Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy  
Tel (+39) 0432 547454 - Fax (+39) 0432 547378  
Mail PEC: info.udine@pec.alifax.com  
[info.udine@alifax.com](mailto:info.udine@alifax.com)