

CERTIFICATE OF ANALYSIS

				ESR_COA	_SI305#2190A.pdf
Product Code		Name	Quantity manufactured	Type of Tubes	
SI 305.300-A	LATE	X CONTROLS	120	Greiner	
	I		i	• i	-
Lot Number		Manufacturing Date	Expiry Date		
2190A		11/05/2021	07/11/2021		

Standard Values				Measured Values					Internal	Volume			
	Transmittance	ESR (mm/h) (min-Val-max)		ESR (mm/h) (min-Val-max)		,	Instrument		Serial Number	ОК	Aggregation	checked	
Level 2	1534	6	-	11	9	9	9	TEST 1 BCL		2173	\mathbf{X}	NO	YES
Level 3	980	15	-	22	19	20	20	TEST1 THL		T13290	X	NO	YES
Level 4	637	56	-	74	66	66	68	ROLLER 20 PN		133	X	NO	YES

Controlled By	Cirio Alberto	Date of Control	12/05/2021
Approved By Manufacturing Manager	Pividori Denis	Date of Approval	12/05/2021

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

Released By Quality

Facchin Fabio

Date of Release 12/05/2021

This is a computer-generated document. No signature is required.

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com