

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh


Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella O:6₁
Catalogue no.	TR 1304
Package size	1 ml
Lot no.	9801118
Expiry date	2022-11
Intended use	Anti-Salmonella O:6 ₁ records all <i>Salmonella</i> strains of the group C ₂ -C ₃ , as long as they contain antigen O:6 (= O:6 ₁) in slide agglutination.
Specificity	No agglutination with the serovars of group C ₂ -C ₃ that contain only the O:8 antigen or the antigen combination O:8,20. The O:6 antigen of the group C ₁ (= O: 6 ₂) is not detected.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin,
place, date

30.01.17


signature Head of Quality Control