

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321  
13088 Berlin, Germany

T: +49 30 927 030-0  
F: +49 30 927 030-30  
info@sifin.de  
www.sifin.de

|                    |  |
|--------------------|--|
| <b>Product</b>     | <b>Anti-Salmonella O:9</b>   |
| Catalogue no.      | TR 1307  |
| Package size       | 1 ml   |
| <b>Lot no.</b>     | <b>25 02 15</b>  |
| <b>Expiry date</b> | <b>2020-02</b>   |
| Intended use       | Anti-Salmonella O:9 records all <i>Salmonella</i> strains, as long as they contain antigen O:9 (group D) in slide agglutination. |
| Specificity        | No agglutination with heterologous <i>Salmonella</i> O antigens and <i>Salmonella</i> H antigens.                                |
| Preservatives      | Sodium azide 0.9 mg/ml   |
| Storage            | 2 °C...8 °C  |

Berlin, 15.05.15  
place, date

  
signature Head of Quality Control