

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 700 144-0
F: +49 30 700 144-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella O:15
Catalogue no.	TR1310
Package size	1 ml
Lot no.	0630120
Expiry date	2023-12
Intended use	Anti-Salmonella O:15 records all <i>Salmonella</i> strains of group E1 after conversion by the phages ϵ 15 or ϵ 15 and ϵ 34. Antigens O:15 or O:15,34 replace antigen O:10. Antigen O:10 is no longer detectable after lysogenisation.
Specificity	No agglutination with heterologous <i>Salmonella</i> O antigens and <i>Salmonella</i> H antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 02.03.2020
place, date


signature Head of Quality Control