

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella O:28
Catalogue no.	TR 1336
Package size	1 ml
Lot no.	8981019
Expiry date	2024-10
Intended use	Anti-Salmonella O:28 records all <i>Salmonella</i> strains, as long as they contain antigen O:28 in slide agglutination
Specificity	No agglutination with heterologous <i>Salmonella</i> O antigens and <i>Salmonella</i> H antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C... 8 °C

Berlin, 13.12.19
place, date

7. Luderj
signature Head of Quality Control