



Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella O:50
Catalogue no.	TR 1355
Package size	1 ml
Lot no.	105 05 14
Expiry date	2019-05
Intended use	Anti-Salmonella O:50 records all <i>Salmonella</i> strains, as long as they contain antigen O:50 in slide agglutination
Specificity	No agglutination with heterologous <i>Salmonella</i> O antigens and <i>Salmonella</i> H antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 11.06.14
place, date


signature Head of Quality Control