

sifin

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Salmonella H:d Test Antigen for the Widal Reaction
Catalogue no.	TS 1603
Package size	10 ml
Lot no.	242 11 15
Expiry date	2016-11
Intended use	The Test antigen reacts in the Widal test with the homologous Salmonella antibody (Anti-H:d) in the serum to be tested. The specific bonding is indicated by agglutination.
Specificity	There must be no reaction with physiological saline solution.
Preservatives	Formalin 0.5 %
Storage	2 °C...8 °C

Berlin, 03.12.15

place, date



signature Head of Quality Control