

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ	N°9072 - 9072/1
	<i>QUALITY CONTROL CERTIFICATE</i>	Revisione 7 del 09.10.2013 Pag. 1 di 1

CARATTERISTICHE DEL PRODOTTO / *PRODUCT CHARACTERISTICS*

Prodotto / <i>Product</i>	ECONAZOLE
Sigla / <i>Initial</i>	ECN 10 ug
Lotto / <i>Batch</i>	060817074
Data produzione / <i>Production date</i>	08.06.2017
Data scadenza / <i>Expiry date</i>	2020.06.07
Contenuto antibiotico del disco / <i>Antimicrobial Disk content</i>	Econazole 10 ug

CONTROLLO QUALITÀ / *QUALITY CONTROL*

Terreno di coltura / <i>Culture medium</i>	Mueller Hinton+Glucose+Methylen Blue Agar	
Tempo di incubazione / <i>Incubation time</i>	20-24 h <i>in caso di crescita insufficiente prolungare l'incubazione fino a 48 ore / read at 48 hours only when insufficient growth is observed after 24 hours of incubation</i>	
Temperatura / <i>Temperature</i>	36 ± 1°C	
Modalità di incubazione / <i>Procedure of incubation</i>	O₂	
Microrganismo/ <i>Microorganism</i>	RESULTS mm	
<i>C. albicans</i>	ATCC 90028	26
<i>C. krusei</i>	ATCC 6258	25
<i>C. parapsilosis</i>	ATCC 22019	25
<i>C. tropicalis</i>	ATCC 750	20

LOTTO / *BATCH* : Idoneo / Approved

Non Idoneo / Not approved

DATA / *DATE* : 15.06.2017

Responsabile Controllo Qualità
Quality Control Manager
(D. Vitagliano)

Dario Vitagliano