



INSTRUCTION MANUAL

REF 3920

April 22, 2020

GA CoV-2 IgG

- 96 determinations -



IVD *In vitro* diagnostic device

Enzyme immunoassay for the determination of IgG antibodies to SARS-Coronavirus 2 in human serum and plasma

REF	Catalogue number	LOT	Batch code
	Consult accompanying documents		Manufactured by
	Temperature limitation		Use by
	Consult operating instruction		Biological risk



GA GENERIC ASSAYS GmbH

Ludwig-Erhard-Ring 3

15827 Dahlewitz, Germany

Telephone: +49 (0) 33708-9286-0

Fax: +49 (0) 33708-9286-50

www.genericassays.com

INTENDED USE

Enzyme Immunoassay for the determination of IgG antibodies to SARS-Coronavirus 2 (SARS-CoV-2) in human serum and plasma for the monitoring of immune response in COVID-19 disease.

Corona Virus Disease 2019 (COVID-19) is caused by the SARS-associated coronavirus 2 (SARS-CoV-2, previously known as "2019-nCoV), which was first identified in a respiratory disease outbreak in Wuhan City, Hubei Province, China. It is since then responsible for a global pandemic. SARS-CoV-2 is a single-stranded RNA virus with positive polarity and belongs to the genus of beta-coronaviruses, which also includes SARS-CoV (2003) and MERS-CoV (2012). Like all other corona viruses, the genome of SARS-CoV-2 encodes the spike protein, the envelope protein, the membrane protein and the nucleocapsid protein.

Patients infected with SARS-CoV-2 may show mild or no symptoms. Common symptoms of COVID-19 are similar to those of a cold or flu. It can take up to 14 days after exposure to SARS-CoV-2 before they appear. Symptoms include: fever, cough, breathing difficulty, pneumonia in both lungs. In severe cases, the infection can lead to death. Current tests for SARS-CoV-2 examine the genetic material of the virus in oral swabs using the polymerase chain reaction (PCR). The PCR only gives a positive result if the virus is still present. The tests cannot identify individuals who have gone through an infection, recovered and have the virus from their bodies removed.

Enzyme immunoassays, on the other hand, are serological tests that allow the determination of specific antibodies to SARS-CoV-2. They are suitable for screening large groups of patients.

1. Wu, A. et al. Genome composition and divergence of the novel coronavirus (2019-nCoV) originating in China. Cell Host Microbe <https://doi.org/10.1016/j.chom.2020.02.001> (2020).
2. Lu, R. et al. Genomic characterization and epidemiology of 2019 novel coronavirus: implications for virus origins and receptor binding. Lancet 395, 565–574 (2020).
3. Zhou, P. et al. Discovery of a novel coronavirus associated with the recent pneumonia outbreak in humans and its potential bat origin. Nature <https://doi.org/10.1038/s41586-020-2012-7> (2020).
4. Zhu, N. et al. A novel coronavirus from patients with pneumonia in China, 2019. N. Engl. J. Med. 382, 727–733 (2020).
5. Huang, C. et al. Clinical features of patients infected with 2019 novel coronavirus in Wuhan, China. Lancet 395, 497–506 (2020).
6. Li, Q. et al. Early transmission dynamics in Wuhan, China, of novel coronavirus-infected pneumonia. N. Engl. J. Med. <https://doi.org/10.1056/nejmoa2001316> (2020).
7. Liu, Y., Gayle, A. A., Wilder-Smith, A. & Rocklöv, J. The reproductive number of COVID-19 is higher compared to SARS coronavirus. J. Travel Med. <https://doi.org/10.1093/jtm/taaa021> (2020).
8. Tang, B. et al. Estimation of the transmission risk of the 2019-nCoV and its implication for public health interventions. J. Clin. Med. 9, 462 (2020).

PRINCIPLE OF THE TEST

GA CoV-2 IgG is used for the determination of IgG antibodies to SARS-Coronavirus-2.

Antibodies of the controls and diluted patient samples react with antigens immobilized on the solid phase of microtiter plates. Use of recombinant antigens guarantees the specific binding of autoimmune antibodies of the specimen under investigation. Following an incubation period of 45 min at 37°C, unbound sample components are removed by a wash step.

The bound IgG antibodies react specifically with anti-human-IgG conjugated to horseradish peroxidase (HRP). Within the incubation period of 45 min at 37°C, excessive conjugate is separated from the solid-phase immune complexes by the following wash step.

HRP converts the added colorless substrate solution of 3,3',5,5'-tetramethylbenzidine (TMB) into a blue product. The enzyme reaction is stopped by dispensing an acidic solution into the wells after 15 min at room temperature (18-25°C) turning the solution from blue to yellow.

The optical density (OD) of the solution measured at 450 nm is directly proportional to the amount of specific antibodies bound.

PATIENT SAMPLES

Specimen collection and storage

Blood is taken by venipuncture. Serum is separated after clotting by centrifugation. Plasma samples (citrate, EDTA, heparin) can be used as well. Hyperlipemic, hemolytic or contaminated samples must not be used and the samples must not contain preservatives.

Samples can be stored up to 5 days at 2 - 8 °C in the primary tubes. For longer storage, sera or plasmas extracted from the primary tubes must be frozen at -20 °C. Repeated freezing and thawing should be avoided. If necessary, aliquots should be prepared before freezing.

Samples are diluted during the test, controls should not be diluted.

TEST COMPONENTS for 96 DETERMINATIONS

A Ag 96	Microtiter plate, 12 breakable strips per 8 wells coated with spike and nucleocapsid antigens of SARS-CoV-2 (recombinant)	1 vacuum sealed with desiccant, 2 adhesive foils
B BUF WASH 20x	Concentrated wash buffer sufficient for 1200 ml solution	60 ml concentrate capped white
C DIL	Sample diluent	50 ml ready for use capped black
G START	Start reagent	8 ml ready for use capped white
D CONJ	Conjugate containing anti-human-IgG coupled with HRP	15 ml ready for use capped red
E SOLN TMB	Substrate 3,3',5,5'-tetramethylbenzidine in citrate buffer containing hydrogen peroxide	15 ml ready for use capped blue
F H2SO4 0.3 M	Stop solution 0.3 M sulfuric acid	15 ml ready for use capped yellow
P CONTROL	Positive control (diluted serum) +	2 ml ready for use capped red
N CONTROL	Negative control (diluted serum) -	2 ml ready for use capped green

Materials required in addition

- calibrated micropipettes (200, 50, 10 µl) with tips
- incubator (37 °C +/- 0.5 °C)
- timer
- 8-channel wash comb with vacuum pump and waste bottle or microplate washer
- microplate reader with optical filters for 450 nm and 620-630 nm (blank)
- absorbent paper
- distilled water
- vortex or similar mixer

Size and storage

GA CoV-2 IgG has been designed for 96 determinations.

The expiry date of the test when stored at 2 – 8 °C is indicated on the outer label. Do not use the test kit after the expiration date.

After opening the test kit, it can be used for 6 months (6 test preparations in total).

Preparation before use

The microtiter plate with divisible strips is sealed in an aluminum-coated pouch together with desiccant. Open the package only after reaching room temperature (approx. 1 h). Do not use the plate if the desiccant bag has turned from yellow to dark green. Protect unused wells from moisture and place them back in the bag together with the desiccant and close it.

The 20-fold concentrated Wash Buffer Solution must be dissolved with distilled water and mixed carefully before use. Carefully dissolve any salt crystals that may be present by warming and shaking before diluting. If necessary, dilute the entire amount of the concentrate. Avoid foaming during preparation, as the presence of bubbles could lead to poor washing efficiency. After dilution the washing solution is stable at +2-8 °C for about 1 week.

All other reagents are ready for use. Please mix well before use (Vortex).

Store substrate protected from light and oxidizing substances.

ASSAY PROCEDURE

Ensure that no fingerprints are present on the bottom of the microtiter plate before OD measurement. Fingerprints could lead to false positive results.

1. Bring all reagents and the required number of test cavities to room temperature (18-25 °C) before use. Mix gently without causing foam.
2. Leave the first well empty for BLANK. Dispense into the respective wells:
200 µl Negative control (N) in triplicate
200 µl Positive control (P)
200 µl Sample diluent (C) into all patient wells.

Add **10 µl of patient samples** to the patient wells. Colour of sample diluent is turning to dark blueish green.
3. Dispense **50 µl of Start reagent (G)** into all wells except BLANK.
4. Cover plate, shake for 30 seconds, incubate **45 Minutes** at 37 °C.
5. Decant, then wash each well **5 times** using **350 µl** wash solution (made of B), use a soak time of 20 seconds each.
6. Add **100 µl** of conjugate (D) solution to all wells except BLANK.
7. Cover plate, incubate **45 min** at 37 °C.
8. Repeat wash step 5.
9. Add **100 µl** of substrate (E) to each well including BLANK.
10. Incubate **15 min protected from light** at room temperature (18...25 °C).
11. Add **100 µl** of stop solution (F) to each well and mix gently.
12. Read the OD at **450 nm** versus 620 (630 nm) within 20 min after adding the stop solution.

If the washing process cannot be carried out with a soak time of the washing buffer, the washing process must be extended by one step.

Pipetting scheme:

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S4										
B	NC	S5										
C	NC	S6										
D	NC	S7										
E	PC	S8										
F	S1	S9										
G	S2	S10										
H	S3	S11										

DATA PROCESSING

The test results are evaluated by calculating a cut-off from the mean value of the Negative Control (N).

$$\text{Cut-off (Co)} = 0.250 + \text{OD}_{\text{Negative control}}$$

The binding index BI is calculated by the ratio of OD values of samples to the Cut-off:

$$\text{BI} = \text{OD}_{\text{sample}} / \text{Co}$$

Interpretation is done according to the following table:

BI	SARS-CoV-2 IgG
< 0.9	negative
0.9 – 1.1	borderline
> 1.1	positive

Patients with borderline results should be retested after a period of 1-2 weeks using a freshly collected sample.

Example of Typical Assay Results

sample	OD	BI
Positive control	1,558	
Negative control (mean)	0,039	
Cut-off	0,289	
Patient 1	1,604	5,5 - positive
Patient 2	0,561	1,9 - positive
Patient 3	0,170	0,6 - negative

Test validity

The test run is valid if:

- OD 450/620 nm of BLANK < 0.100
- OD 450/620 nm of Negative control (mean) < 0.150 after BLANK subtraction
- OD 450/620 nm of Positive control > 0.500.

If the above mentioned quality criteria are not met, repeat the test and make sure that the test procedure is followed correctly (incubation times and temperatures, sample and wash buffer dilution, wash steps etc.). In case of repeated failure of the quality criteria contact your supplier.

PERFORMANCE CHARACTERISTICS

Due to the difficulty of evaluating samples from infected hospital patients, the performance data are still being evaluated.

Precision

Using a negative and a weakly positive sample, the precision profiles were created in 16 repetitions in three separate runs. The variability did not lead to an incorrect sample classification.

Sensitivity

A first study conducted at an emergency center using samples from a cohort of infected patients showed a sensitivity of about 98%. Further extensive studies are still being evaluated.

Specificity

The test has been performed on hundreds of samples collected before the COVID-19 outbreak. A specificity of > 95% was determined. In a small number of the "normal" population IgG antibodies to the nucleocapsid antigen have been found. These antibodies are probably formed during previous contact with other members of the coronavirus family.

Limitations of Method

Performing the test with lipemic or hemolytic samples may result in false positive results.

Any clinical diagnosis should not be based on the results of in vitro diagnostic methods alone. Physicians are supposed to consider all clinical and laboratory findings possible to state a diagnosis.

A positive result should be confirmed with a test method for the determination of IgG antibodies to single antigens like GA CoV IgG+.

Patients may be negative for IgG antibodies during the first week of infection with SARS-CoV-2.

Negative results may also occur in patients with immunodeficiency, diseases that affect immune function, when taking medication to suppress immune function, or in the failure of major systemic organs.

Previous infections with SARS or another coronavirus strain may cause a mild IgG response, given the similarity of the different strains.

INCUBATION SCHEME

GA CoV-2 IgG (3920)

		Controls (P, N)	Patient samples	Blank
1.	Bring all test reagents to room temperature (18...25 °C).			
2.	Dispense	Controls (P/N)	200 µl	
		Sample diluent (C)		200 µl
		Patient samples		10 µl
3.	Dispense	Start reagent (G)	50 µl	50 µl
4.	Incubate	Shake for 30 sec, 45 min at 37 °C		
5.	Wash	5 x 350 µl (soak time 20 sec)		
6.	Dispense	Conjugate (D)	100 µl	100 µl
7.	Incubate	45 min at 37 °C		
8.	Wash	5 x 350 µl (soak time 20 sec)		
9.	Dispense	Substrate (E)	100 µl	100 µl
10.	Incubate	15 min at room temperature (18-25 °C) in the dark		
11.	Dispense	Stop solution (F)	100 µl	100 µl
12.	Read	at 450 nm against 620 (630) nm		

SAFETY PRECAUTIONS

- **This kit is for in vitro use only.** Follow the working instructions carefully. This instruction manual is valid only for the present kit with the given composition. The exchange of individual components is only permitted within the framework of a compatibility list of the manufacturer.
- The kit should be performed by trained technical staff only.
- The test kit or its opened reagents should only be used within the specified shelf life.
- All reagents should be kept at 2 - 8 °C in the original package until use.
- Do not use or mix reagents from different lots. Do not use reagents from other manufacturers.
- Some of the reagents contain small amounts of ProClin 300 (< 1.0 % v/v) und Natriumazid (< 0.1%) as preservative. They must not be swallowed or allowed to come into contact with skin or mucosa.
- Source materials derived from human body fluids or organs used in the preparation of this kit were tested and found negative for HBsAg and HIV as well as HCV antibodies. However, no known test guarantees the absence of such viral agents. Therefore, handle all components and all patient samples as if potentially hazardous.
- Since the kit contains potentially hazardous materials, the following precautions should be observed:
 - Do not smoke, eat or drink while handling kit material,
 - Always use protective gloves,
 - Never pipette material by mouth,
 - Wipe up spills promptly, washing the affected surface thoroughly with a decontaminant.
- In any case GLP should be applied with all general and individual regulations to the use of this kit.