

<b>Liofilchem</b> S.r.l.	CERTIFICATO CONTROLLO QUALITÀ	N°9072 - 9072/1
	<i>QUALITY CONTROL CERTIFICATE</i>	Revisione 7 del 09.10.2013 Pag. 1 di 1

### CARATTERISTICHE DEL PRODOTTO / *PRODUCT CHARACTERISTICS*

Prodotto / <i>Product</i>	<b>ECONAZOLE</b>
Sigla / <i>Initial</i>	<b>ECN 10 ug</b>
Lotto / <i>Batch</i>	<b>073119138</b>
Data produzione / <i>Production date</i>	<b>31.07.2019</b>
Data scadenza / <i>Expiry date</i>	<b>2022.07.30</b>
Contenuto antibiotico del disco / <i>Antimicrobial Disk content</i>	<b>Econazole 10 ug</b>

### CONTROLLO QUALITÀ / *QUALITY CONTROL*

Terreno di coltura / <i>Culture medium</i>	<b>Mueller Hinton+Glucose+Methylen Blue Agar</b>	
Tempo di incubazione / <i>Incubation time</i>	<b>20-24 h</b> <i>in caso di crescita insufficiente prolungare l'incubazione fino a 48 ore / read at 48 hours only when insufficient growth is observed after 24 hours of incubation</i>	
Temperatura / <i>Temperature</i>	<b>36 ± 1°C</b>	
Modalità di incubazione / <i>Procedure of incubation</i>	<b>O<sub>2</sub></b>	
Microrganismo/ <i>Microorganism</i>	<b>RESULTS</b> mm	
<i>C. albicans</i>	ATCC 90028	34
<i>C. krusei</i>	ATCC 6258	33
<i>C. parapsilosis</i>	ATCC 22019	34
<i>C. tropicalis</i>	ATCC 750	28

LOTTO / *BATCH* :  Idoneo / Approved

Non Idoneo / Not approved

DATA / *DATE* : 07.08.2019

Responsabile Controllo Qualità  
*Quality Control Manager*  
(D. Vitagliano)

*Dario Vitagliano*